



We care about your needs

Halotech DNA SL.
C/ Faraday 7, 28049 Madrid
Spain
Tel: +34 91 279 6950
info@halotech.es
www.halotechdna.com

CERTIFICATE OF ANALYSIS

Product name: Halosperm®G2
Product number: HT-HSG2
Lot Number: G2 1726

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Lysis Solution, pH 7,0±1, Colorless, Transparent, Aliquoted volume 10 ml.

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Denaturant Agent, pH 1,0±1, Smell HCl, Translucent appearance Transparent, Aliquoted volume 10 ml.

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Supercoated slides, Translucent appearance Transparent, Homogeneity surface completely covered.

Table with 4 columns: COMPONENTS, Parameter/Test, Specification, Result. Row 1: Eppendor tube with Agarose, Color Colorless, State Solid, Aliquoted volume 1,5 ml.

Results with biological sample, comparison two tests: Accepted/Rejected (AC/RE) AC

Table with 2 columns: Expiration information, The expiry date is guaranteed to be valid for twelve months from the production date provided. Production Date 07/11/2017, Expiration Date 07/11/2018, Certificate Issue date 08/11/2017.

Handwritten signature of Francisco Arenal

Francisco Arenal
Quality Assurance Manager



NOTICE: HalotechDNA products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from HalotechDNA.