



We care about your needs

Halotech DNA SL.
C/ Faraday 7 , 28049 Madrid
Spain
Tel: +34 91 279 6950
info@halotech.es
www.halotechdna.com

CERTIFICATE OF ANALYSIS

Product name: Halosperm®G2

Product number: HT-HSG2

Lot Number: G2 1818

COMPONENTS	Parameter/Test	Specification	Result
Lysis Solution	pH	7,0±1	7,0
	Color	Colorless	Colorless
	Translucent appearance	Transparent	Transparent
	Aliquoted volume	10 ml	10 ml

Denaturant Agent	pH	1,0±1	1,0
	Smell	HCl	HCl
	Translucent appearance	Transparent	Transparent
	Aliquoted volume	10 ml	10 ml

Supercoated slides	Translucent appearance	Transparent	Transparent
	Homogeneity	surface completely covered	surface completely covered

Eppendor tube with Agarose	Color	Colorless	Colorless
	State	Solid	Solid
	Aliquoted volume	1,5 ml	1,5 ml

Results with biological sample, comparison two tests:

Accepted/Rejected (AC/RE)

AC

Expiration information	The expiry date is guaranteed to be valid for twelve months from the production date provided.		
Production Date	29/05/2018		
Expiration Date	29/05/2019		
Certificate Issue date	31/05/2018		

Francisco Arenal
Quality Assurance Manager



NOTICE: HalotechDNA products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from HalotechDNA.