



We care about your needs

Halotech DNA SL.
C/ Faraday 7 , 28049 Madrid
Spain
Tel: +34 91 279 6950
info@halotech.es
www.halotechdna.com

CERTIFICATE OF ANALYSIS

Product name: oxiSperm®
Product number: HT-OS20
Lot Number: OS20 1802

COMPONENTS	Parameter/Test	Specification	Result
Reactive Gel	pH	7,0±1	7,1
	Color	Yellowish	Yellowish
	Translucent appearance	Transparent	Transparent
	Aliquoted volume	1 ml	1 ml

Results with biological sample, comparison two tests:

Accepted/Rejected (AC/RE)

AC

Expiration information	The expiry date is guaranteed to be valid for twelve months from the production date provided.
Production Date	13/04/2018
Expiration Date	13/04/2019
Certificate Issue date	16/04/2018

Francisco Arenal
Quality Assurance Manager



NOTICE: HalotechDNA products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from HalotechDNA.