

## Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP16EP02
Production date:	DD/MM/YYYY	01/03/16
Expiry date:	DD/MM/YYYY	31/03/18
Batch size:		100 kits

	<i>Result</i>	<i>Specification</i>
--	---------------	----------------------

**Reagent 1**

Osmolality (mOsm/kg):	229	210-260 mOsm/kg
pH:	6,76	6,70 - 6,90

**Reagent 3**

Osmolality (mOsm/kg):	1028	930-1140 mOsm/kg
pH:	6,55	6,55 - 6,75

**Reagent 4**

pH:	12,07	> 12
-----	-------	------

**Reagent 6**

pH:	12,17	> 12
-----	-------	------

Functionality test:	PASSED	PASSED
---------------------	--------	--------

Product released by Justine Van Renterghem on 18/03/2016 (DD/MM/YYYY)

