

Certificate of Analysis and Release

FertiCult Aspiration medium

Batchnumber:		FP16FA04
Production date:	DD/MM/YYYY	15/03/2016
Expiry date:	DD/MM/YYYY	30/09/2017
Total volume:		70000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,41	7.30-7.60
Osmolality:	277	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml

Mouse embryo test (1-cell):	95%	≥ 80%
-----------------------------	-----	-------

Product released by Lenie Bocklandt on 30/03/2016 (DD/MM/YYYY)



Backlandt lenie