

Certificate of Analysis and Release

FertiCult Aspiration medium

Batchnumber:		FP16FA09
Production date:	DD/MM/YYYY	07/06/2016
Expiry date:	DD/MM/YYYY	31/12/2017
Total volume:		80000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,40	7.30-7.60
Osmolality:	277	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/mL	< 0,25EU/ml
Mouse embryo test (1-cell):	100%	≥ 80%

Product released by Lenie Bocklandt on 22/06/2016 (DD/MM/YYYY)



Backlandt Lenie