

Certificate of Analysis and Release

FertiCult IVF Medium

Batchnumber:		FP16FC06
Production date:	DD/MM/YYYY	31/03/2016
Expiry date:	DD/MM/YYYY	31/03/2017
Total volume:		25000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,40	7.20-7.50
Osmolality:	280	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml
Mouse embryo test (1-cell):	80,95%	≥ 80%

Product released by Justine Van Renterghem on 15/04/2016 (DD/MM/YYYY)

