

## Certificate of Analysis and Release

### FertiCult IVF Medium

Batchnumber:		FP16FC09
Production date:	DD/MM/YYYY	01/06/2016
Expiry date:	DD/MM/YYYY	30/06/2017
Total volume:		30000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,42	7.20-7.50
Osmolality:	274	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml
Mouse embryo test (1-cell):	90,48 %	≥ 80%

Product released by Justine Van Renterghem on 16/06/2016 (DD/MM/YYYY)

