

## Certificate of Analysis and Release

### FertiCult Flushing Medium

Batchnumber:		FP16FL01PRG
Production date:	DD/MM/YYYY	05/01/2016
Expiry date:	DD/MM/YYYY	31/07/2017
Total volume:		20000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,40	7.30-7.60
Osmolality:	280	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	<0,05 EU/ml	< 0,25EU/ml

Mouse embryo test (1-cell):	100%	≥ 80%
-----------------------------	------	-------

Product released by Lenie Bocklandt on 20/01/2016 (DD/MM/YYYY)



Backlandt lenie