

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP16FL04
Production date:	DD/MM/YYYY	28/01/2016
Expiry date:	DD/MM/YYYY	31/07/2017
Total volume:		44000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,47	7.30-7.60
Osmolality:	279	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml

Mouse embryo test (1-cell):	100%	≥ 80%
-----------------------------	------	-------

Product released by Justine Van Renterghem on 16/02/2016 (DD/MM/YYYY)

