

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP16FL08
Production date:	DD/MM/YYYY	18/02/2016
Expiry date:	DD/MM/YYYY	31/08/2017
Total volume:		50000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,36	7.30-7.60
Osmolality:	282	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml

Mouse embryo test (1-cell):	95,24%	≥ 80%
-----------------------------	--------	-------

Product released by Justine Van Renterghem on 04/03/2016 (DD/MM/YYYY)

