

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP16FL14
Production date:	DD/MM/YYYY	14/03/2016
Expiry date:	DD/MM/YYYY	30/09/2017
Total volume:		10000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,33	7.30-7.60
Osmolality:	281	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml

Mouse embryo test (1-cell):	86%	≥ 80%
-----------------------------	-----	-------

Product released by Lenie Bocklandt on 30/03/2016 (DD/MM/YYYY)



Backlandt Lenie

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.