

## Certificate of Analysis and Release

Product: Hypo-osmotic swelling test

Batchnumber :		FP16HOS02
Production date :	DD/MM/YYYY	18/05/2016
Expiry date :	DD/MM/YYYY	31/05/2017
Batch size:		6000 ml

	<i>Result</i>	<i>Specification</i>
Osmolality:	167 mOsm/kg	160-170mOsm/kg
pH:	7,22	7.2-7.6
Functionality test:	PASSED	PASSED

Product released by Justine Van Renterghem on 02/06/2016 (DD/MM/YYYY)

