

## Certificate of Analysis and Release

Product: Hyaluronidase in Flushing medium

Batchnumber :		FP16HY04
Production date :	DD/MM/YYYY	8/03/2016
Expiry date :	DD/MM/YYYY	31/03/2017
Batch size:		310 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	279 mOsm/kg	270-290mOsm/kg
pH:	7,54	7.3-7.6
Endotoxin:	< 0,2 EU/ml	< 1.00EU/ml
Sterility test:	PASSED	PASSED
Mouse embryo test:	100%	>=80%

Product released by Lenie Bocklandt on 23/03/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to read "Lenie Bocklandt".

Bocklandt lenie