

## Certificate of Analysis and Release

Product: Hyaluronidase in Flushing medium

Batchnumber :		FP16HY05
Production date :	DD/MM/YYYY	16/03/2016
Expiry date :	DD/MM/YYYY	31/03/2017
Batch size:		166 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	280 mOsm/kg	270-290mOsm/kg
pH:	7,38	7.3-7.6
Endotoxin:	< 0,2 EU/ml	< 1.00EU/ml
Sterility test:	PASSED	PASSED
Mouse embryo test:	81%	>=80%

Product released by Justine Van Renterghem on 31/03/2016 (DD/MM/YYYY)

