

Certificate of Analysis and Release

Product: Hyaluronidase in Flushing medium

Batchnumber :		FP16HY08
Production date :	DD/MM/YYYY	22/04/2016
Expiry date :	DD/MM/YYYY	30/04/2017
Batch size:		217 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	278 mOsm/kg	270-290mOsm/kg
pH:	7,43	7.3-7.6
Endotoxin:	< 0,2 EU/ml	< 1.00EU/ml
Sterility test:	PASSED	PASSED
Mouse embryo test:	100%	>=80%

Product released by Justine Van Renterghem on 09/05/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.