

## Certificate of Analysis and Release

Product: 10% PVP solution

Batchnumber :		FP16PVP02
Production date :	DD/MM/YYYY	16-feb-2016
Expiry date :	DD/MM/YYYY	30-nov-2016
Batch size:		50 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	293 mOsm/kg	280-320mOsm/kg
Mouse Embryo Assay (blasts after 96h culture):	90%	>= 80%
pH:	7,39	7.2-7.6
Endotoxin:	0,06 EU/ml	< 1EU/ml
Viscosity:	177,6	>170cP
Sterility (fungal and bacterial sterility test):	PASSED	PASSED

Product released by Justine Van Renterghem on 07/03/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.