

## Certificate of Analysis and Release

Product: 10% PVP solution

Batchnumber :		FP16PVP04
Production date :	DD/MM/YYYY	23-mrt-2016
Expiry date :	DD/MM/YYYY	31-dec-2016
Batch size:		197 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	292 mOsm/kg	280-320mOsm/kg
Mouse Embryo Assay (blasts after 96h culture):	100%	>= 80%
pH:	7,33	7.2-7.6
Endotoxin:	< 0,05 EU/ml	< 1EU/ml
Viscosity:	205,2	>170cP
Sterility (fungal and bacterial sterility test):	PASSED	PASSED

Product released by Justine Van Renterghem on 07/04/2016 (DD/MM/YYYY)

