

Certificate of Analysis and Release

Product: 10% PVP solution

Batchnumber :		FP16PVP06
Production date :	DD/MM/YYYY	02-mei-2016
Expiry date :	DD/MM/YYYY	28-feb-2017
Batch size:		163 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	294 mOsm/kg	280-320mOsm/kg
Mouse Embryo Assay (blasts after 96h culture):	100%	>= 80%
pH:	7,31	7.2-7.6
Endotoxin:	< 0,05 EU/ml	< 1EU/ml
Viscosity:	235,1	>170cP
Sterility (fungal and bacterial sterility test):	PASSED	PASSED

Product released by Justine Van Renterghem on 17/05/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.