

Certificate of Analysis and Release

Product: 10% PVP solution

Batchnumber :		FP16PVP08
Production date :	DD/MM/YYYY	20-jun-2016
Expiry date :	DD/MM/YYYY	31-mrt-2017
Batch size:		641 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	295 mOsm/kg	280-320mOsm/kg
Mouse Embryo Assay (blasts after 96h culture):	95,24%	>= 80%
pH:	7,34	7.2-7.6
Endotoxin:	0,2 EU/ml	< 1EU/ml
Viscosity:	193,2	>170cP
Sterility (fungal and bacterial sterility test):	PASSED	PASSED

Product released by Lenie Bocklandt on 05/07/2016 (DD/MM/YYYY)



Bocklandt lenie