

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP16SF03
Production date:	DD/MM/YYYY	9/02/2016
Expiry date:	DD/MM/YYYY	31/08/2017
Total number of kits:		1279 bottles

	Results	Specification
pH:	7,38	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (45' min exposure):	92%	≥ 80%
Endotoxin:	< 0,05 EU/ml	< 0,25 EU/mL

Product released by Lenie Bocklandt on 24/02/2016 (DD/MM/YYYY)



Bocklandt lenie

All blood products should be treated as potentially infectious. Source materials from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and HTLV I/II and non-reactive for HBsAg, HCV RNA, and HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents.