

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP16SF05
Production date:	DD/MM/YYYY	10/03/2016
Expiry date:	DD/MM/YYYY	30/09/2017
Total number of kits:		186 bottles

	Results	Specification
pH:	7,47	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (45' min exposure):	99%	≥ 80%
Endotoxin:	< 0,05 EU/ml	< 0,25 EU/mL

Product released by Justine Van Renterghem on 25/03/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be 'Justine Van Renterghem', written over a horizontal line.