

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP16SF06
Production date:	DD/MM/YYYY	1/04/2016
Expiry date:	DD/MM/YYYY	31/10/2017
Total number of kits:		1000 bottles

	Results	Specification
pH:	7,47	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (45' min exposure):	99%	≥ 80%
Endotoxin:	< 0,05 EU/ml	< 0,25 EU/mL

Product released by Justine Van Renterghem on 18/04/2016 (DD/MM/YYYY)

