

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP16SF07
Production date:	DD/MM/YYYY	14/04/2016
Expiry date:	DD/MM/YYYY	31/10/2017
Total number of kits:		826 bottles

	Results	Specification
pH:	7,50	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (45' min exposure):	91%	≥ 80%
Endotoxin:	< 0,05 EU/ml	< 0,25 EU/mL

Product released by Justine Van Renterghem on 28/04/2016 (DD/MM/YYYY)

