

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP16SF08
Production date:	DD/MM/YYYY	9/05/2016
Expiry date:	DD/MM/YYYY	30/11/2017
Total number of kits:		368 bottles

	Results	Specification
pH:	7,37	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (4 hours exposure):	95%	≥ 80%
Endotoxin:	< 0,0500 EU/ml	< 0,25 EU/mL

Product released by Lenie Bocklandt on 24/05/2016 (DD/MM/YYYY)



Bocklandt lenie