

## Certificate of Analysis and Release

### Product: SpermFreeze

Batchnumber:		FP16SF10
Production date:	DD/MM/YYYY	16/06/2016
Expiry date:	DD/MM/YYYY	31/12/2017
Total number of kits:		658 bottles

	Results	Specification
pH:	7,48	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (4 hours exposure):	98%	≥ 80%
Endotoxin:	< 0,05 EU/ml	< 0,25 EU/mL

Product released by Justine Van Renterghem on 01/07/2016 (DD/MM/YYYY)

