

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP16SIP01U
Production date :	DD/MM/YYYY	11/01/16
Expiry date :	DD/MM/YYYY	31/07/17
Total volume :		23000 ml

Sil-select Plus: Upper layer

	<i>Result</i>	<i>Specification</i>
pH :	7,33	7.20 - 7.60
Osmolality :	320	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL

Product released by Justine Van Renterghem on 26/01/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to read 'Justine Van Renterghem', is written over a horizontal line.