

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

| | | |
|-------------------|------------|------------|
| Batchnumber : | | FP16SIP03U |
| Production date : | DD/MM/YYYY | 2/02/2016 |
| Expiry date : | DD/MM/YYYY | 31/08/2017 |
| Total volume : | | 25625 ml |

| Sil-select Plus: Upper layer | <i>Result</i> | <i>Specification</i> |
|-------------------------------------|---------------|----------------------|
| pH : | 7,33 | 7.20 - 7.60 |
| Osmolality : | 322 | 310-340mOsm/kg |
| Bacterial screen : | NEGATIVE | NEGATIVE |
| Fungal screen : | NEGATIVE | NEGATIVE |
| Endotoxin : | < 0,1 EU/ml | < 0,5EU/mL |

Product released by Lenie Bocklandt on 17/02/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Lenie Bocklandt".

Bocklandt denie