

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP16SIP04L
Production date :	DD/MM/YYYY	11/02/2016
Expiry date :	DD/MM/YYYY	31/08/2017
Total volume :		46200 ml

Sil-select Plus: Lower layer

	<i>Result</i>	<i>Specification</i>
pH :	7,38	7.20 - 7.60
Osmolality :	331	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0.1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1073	1.1050 - 1.1150 g/mL
Sperm survival test:	97%	≥ 80%
Viscosity in cP (25 °C):	1,52	< 1.65 cP

Product released by Lenie Bocklandt on 29/02/2016 (DD/MM/YYYY)



Bocklandt denie