

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP16SIP04U
Production date :	DD/MM/YYYY	11/02/2016
Expiry date :	DD/MM/YYYY	31/08/2017
Total volume :		13000 ml

Sil-select Plus: Upper layer

	<i>Result</i>	<i>Specification</i>
pH :	7,34	7.20 - 7.60
Osmolality :	310	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL

Product released by Justine Van Renterghem on 26/02/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.