

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP16SIP05U
Production date :	DD/MM/YYYY	18/02/2016
Expiry date :	DD/MM/YYYY	31/08/2017
Total volume :		14750 ml

Sil-select Plus: Upper layer

	<i>Result</i>	<i>Specification</i>
pH :	7,39	7.20 - 7.60
Osmolality :	319	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL

Product released by Justine Van Renterghem on 04/03/2016 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be 'Justine Van Renterghem', written over a horizontal line.