

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP16SIP08U
Production date :	DD/MM/YYYY	6/04/2016
Expiry date :	DD/MM/YYYY	31/10/2017
Total volume :		16500 ml

<i>Sil-select Plus: Upper layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,31	7.20 - 7.60
Osmolality :	313	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL

Product released by Justine Van Renterghem on 21/04/2016 (DD/MM/YYYY)

