

## Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP16SIP14L
Production date :	DD/MM/YYYY	31/05/2016
Expiry date :	DD/MM/YYYY	30/11/2017
Total volume :		45575 ml

<b>Sil-select Plus: Lower layer</b>	<i>Result</i>	<i>Specification</i>
pH :	7,29	7.20 - 7.60
Osmolality :	328	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1058	1.1050 - 1.1150 g/mL
Sperm survival test:	97 %	≥ 80%
Viscosity in cP (25 °C):	1,47	< 1.65 cP

Product released by Lenie Bocklandt on 15/06/2016 (DD/MM/YYYY)



Bocklandt denie