

## Certificate of Analysis and Release

Product: Sil-Select Stock

Batchnumber :		FP16SIS02
Production date :	DD/MM/YYYY	1/02/2016
Expiry date :	DD/MM/YYYY	31/08/2017
Total volume :		80000 ml

	<i>Result</i>	<i>Specification</i>
pH :	7,39	7.20 - 7.60
Osmolality :	312	300 - 330 mOsm/kg
Sterility test bacteria:	NEGATIVE	NEGATIVE
Sterility test fungi:	NEGATIVE	NEGATIVE
Endotoxin:	< 0,1 EU/ml	< 0,5 EU/mL
Sperm survival test :	98%	≥ 80%
Density (g/mL):	1,1230	1,1150-1,1250 g/mL
Viscosity in cP (25°C):	1,56	< 1,75 cP

Product released by Justine Van Renterghem on 16/02/2016 (DD/MM/YYYY)

