

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP16VTE03
Production date :	DD/MM/YYYY	17/02/2016
Expiry date :	DD/MM/YYYY	28/02/2017
Total batchsize:		221 Kits

VitriThaw ES medium 5

	Result	Specification
pH :	7,20	7.2 - 7.4
Osmolality:	283	270-290mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxine :	< 0,1 EU/mL	< 0,25EU/mL

VitriThaw ES medium 4

pH :	7,21	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Osmolality :	418	405-435mOsm/kg

VitriThaw ES medium 3

pH :	7,20	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Osmolality :	559	535-565mOsm/kg

VitriThaw ES medium 2

pH :	7,20	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/mL	< 0,25EU/mL
Osmolality :	845	805-850mOsm/kg

VitriThaw ES medium 1

pH :	7,24	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/mL	< 0,25EU/mL

Recovery assay performance on 1-cell mouse embryos: 100% ≥ 80% blastocyst after 96h

Product released by Justine Van Renterghem on 07/03/2016 (DD/MM/YYYY)



Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.