

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP16VTE05
Production date :	DD/MM/YYYY	29/02/2016
Expiry date :	DD/MM/YYYY	28/02/2017
Total batchsize:		188 Kits

VitriThaw ES medium 5	Result	Specification
pH :	7,21	7.2 - 7.4
Osmolality:	285	270-290mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxine :	< 0,1 EU/mL	< 0,25EU/mL

VitriThaw ES medium 4		
pH :	7,28	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1 EU/mL	< 0,25EU/mL
Osmolality :	426	405-435mOsm/kg

VitriThaw ES medium 3		
pH :	7,31	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/mL	< 0,25EU/mL
Osmolality :	556	535-565mOsm/kg

VitriThaw ES medium 2		
pH :	7,22	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/mL	< 0,25EU/mL
Osmolality :	830	805-850mOsm/kg

VitriThaw ES medium 1		
pH :	7,20	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/mL	< 0,25EU/mL

Recovery assay performance on 1-cell mouse embryos: 95% ≥ 80% blastocyst after 96h

Product released by Justine Van Renterghem on 15/03/2016 (DD/MM/YYYY)



Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.