

Certificate of Analysis and Release

Product: EpiScreen Plus

| | | |
|------------------|------------|------------|
| Batchnumber: | | FP17EP03 |
| Production date: | DD/MM/YYYY | 26/04/2017 |
| Expiry date: | DD/MM/YYYY | 30/04/2019 |
| Batch size: | | 100 kits |

| | <i>Result</i> | <i>Specification</i> |
|-----------------------|---------------|----------------------|
| Reagent 1 | | |
| Osmolality (mOsm/kg): | 232 | 210-260 mOsm/kg |
| pH: | 6,76 | 6,70 - 6,90 |
| Reagent 3 | | |
| Osmolality (mOsm/kg): | 1012 | 930-1140 mOsm/kg |
| pH: | 6,59 | 6,55 - 6,75 |
| Reagent 4 | | |
| pH: | 12,16 | > 12 |
| Reagent 6 | | |
| pH: | 12,26 | > 12 |
| Functionality test: | PASSED | PASSED |

Product released by Justine Van Renterghem on 10/05/2017 (DD/MM/YYYY)

