

## Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP17EP04
Production date:	DD/MM/YYYY	23/05/2017
Expiry date:	DD/MM/YYYY	31/05/2019
Batch size:		100 kits

	<i>Result</i>	<i>Specification</i>
--	---------------	----------------------

**Reagent 1**

Osmolality (mOsm/kg):	236	210-260 mOsm/kg
pH:	6,73	6,70 - 6,90

**Reagent 3**

Osmolality (mOsm/kg):	1028	930-1140 mOsm/kg
pH:	6,57	6,55 - 6,75

**Reagent 4**

pH:	12,21	> 12
-----	-------	------

**Reagent 6**

pH:	12,28	> 12
-----	-------	------

Functionality test:	PASSED	PASSED
---------------------	--------	--------

Product released by Justine Van Renterghem on 21/06/2017 (DD/MM/YYYY)

