

Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP17EP06
Production date:	DD/MM/YYYY	11/10/2017
Expiry date:	DD/MM/YYYY	31/10/2019
Batch size:		99 kits

	<i>Result</i>	<i>Specification</i>
Reagent 1		
Osmolality (mOsm/kg):	228	210-260 mOsm/kg
pH:	6,80	6,70 - 6,90
Reagent 3		
Osmolality (mOsm/kg):	1014	930-1140 mOsm/kg
pH:	6,64	6,55 - 6,75
Reagent 4		
pH:	12,15	> 12
Reagent 6		
pH:	12,20	> 12
Functionality test:	PASSED	PASSED

Product released by Eileen Vanderplancke on 07/11/2017 (DD/MM/YYYY)

Eileen Vanderplancke