

Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP17EP07
Production date:	DD/MM/YYYY	18/12/2017
Expiry date:	DD/MM/YYYY	31/12/2019
Batch size:		100 kits

	<i>Result</i>	<i>Specification</i>
--	---------------	----------------------

Reagent 1

Osmolality (mOsm/kg):	233	210-260 mOsm/kg
pH:	6,81	6,70 - 6,90

Reagent 3

Osmolality (mOsm/kg):	1042	930-1140 mOsm/kg
pH:	6,65	6,55 - 6,75

Reagent 4

pH:	12,10	> 12
-----	-------	------

Reagent 6

pH:	12,07	> 12
-----	-------	------

Functionality test:	PASSED	PASSED
---------------------	--------	--------

Product released by Lenie Bocklandt on 04/01/2018 (DD/MM/YYYY)



Bocklandt lenie