

## Certificate of Analysis and Release

### FertiCult Aspiration medium

Batchnumber:		FP17FA08
Production date:	DD/MM/YYYY	11/05/2017
Expiry date:	DD/MM/YYYY	30/11/2018
Total volume:		80000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,44	7.30-7.60
Osmolality:	284	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	0,0915 EU/mL	< 0,25EU/ml
Concentration heparin:	2,997	2 - 3.75 g/liter
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Eileen Vanderplancke on 26/05/2017 (DD/MM/YYYY)

*Eileen Vanderplancke*