

Certificate of Analysis and Release

FertiCult Aspiration medium

| | | |
|------------------|------------|------------|
| Batchnumber: | | FP17FA12 |
| Production date: | DD/MM/YYYY | 31/07/2017 |
| Expiry date: | DD/MM/YYYY | 31/01/2019 |
| Total volume: | | 80000 ml |

| | <i>Results</i> | <i>Specifications</i> |
|-----------------------------|----------------|-----------------------|
| pH: | 7,46 | 7.30-7.60 |
| Osmolality: | 278 | 270-290mOsm/kg |
| Bacterial screen: | PASSED | PASSED |
| Fungal screen: | PASSED | PASSED |
| Endotoxin: | < 0,05 EU/mL | < 0,25EU/ml |
| Concentration heparin: | 2,394 | 2 - 3.75 g/liter |
| Mouse embryo test (1-cell): | 100 % | ≥ 80% |

Product released by Eileen Vanderplancke on 16/08/2017 (DD/MM/YYYY)

Eileen Vanderplancke