

Certificate of Analysis and Release

FertiCult Aspiration medium

Batchnumber:		FP17FA19
Production date:	DD/MM/YYYY	07/11/2017
Expiry date:	DD/MM/YYYY	31/05/2019
Total volume:		80000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,40	7.30-7.60
Osmolality:	277	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/mL	< 0,25EU/ml
Concentration heparin:	2,494	2 - 3.75 g/liter
Mouse embryo test (1-cell):	95,24 %	≥ 80%

Product released by Eileen Vanderplancke on 22/11/2017 (DD/MM/YYYY)

Eileen Vanderplancke

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.