

Certificate of Analysis and Release

FertiCult IVF Medium

Batchnumber:		FP17FC03PRG
Production date:	DD/MM/YYYY	07/02/2017
Expiry date:	DD/MM/YYYY	28/02/2018
Total volume:		15000 ml

	<i>Results</i>	<i>Specifications</i>
pH (@37°C and 5-6% CO ₂):	7,43	7.20-7.50
Osmolality:	281	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	4,073 mg/mL	3.2 - 4.8 g/liter
Concentration gentamicin:	7,94 mg/L	5 - 15 mg/liter
Mouse embryo test (1-cell):	100%	≥ 80%

Product released by Justine Van Renterghem on 22/02/2017 (DD/MM/YYYY)

