

Certificate of Analysis and Release

FertiCult IVF Medium

Batchnumber:		FP17FC07
Production date:	DD/MM/YYYY	12/04/2017
Expiry date:	DD/MM/YYYY	30/04/2018
Total volume:		40000 ml

	<i>Results</i>	<i>Specifications</i>
pH (@37°C and 5-6% CO ₂):	7,42	7.20-7.50
Osmolality:	279	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	3,864	3.2 - 4.8 g/liter
Mouse embryo test (1-cell):	95%	≥ 80%

Product released by Justine Van Renterghem on 27/04/2017 (DD/MM/YYYY)

