

## Certificate of Analysis and Release

### FertiCult IVF Medium

Batchnumber:		FP17FC12
Production date:	DD/MM/YYYY	24/07/2017
Expiry date:	DD/MM/YYYY	31/07/2018
Total volume:		40000 ml

	<i>Results</i>	<i>Specifications</i>
pH (@37°C and 5-6% CO <sub>2</sub> ):	7,38	7.20-7.50
Osmolality:	281	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/mL	< 0,25EU/ml
Concentration HSA:	3,777	3.2 - 4.8 g/liter
Mouse embryo test (1-cell):	85,71 %	≥ 80%

Product released by Eileen Vanderplancke on 08/08/2017 (DD/MM/YYYY)

*Eileen Vanderplancke*