

Certificate of Analysis and Release

FertiCult IVF Medium

Batchnumber:		FP17FC17
Production date:	DD/MM/YYYY	29/11/2017
Expiry date:	DD/MM/YYYY	30/11/2018
Total volume:		30000 ml

	<i>Results</i>	<i>Specifications</i>
pH (@37°C and 5-6% CO ₂):	7,46	7.20-7.50
Osmolality:	277	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	<0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	4,01	3.2 - 4.8 g/liter
Mouse embryo test (1-cell):	90%	≥ 80%

Product released by Justine Van Renterghem on 27/12/2017 (DD/MM/YYYY)



Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.