

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP17FL09
Production date:	DD/MM/YYYY	27/02/2017
Expiry date:	DD/MM/YYYY	31/08/2018
Total volume:		135000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,38	7.30-7.60
Osmolality:	279	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05EU/ml	< 0,25EU/ml
Concentration HSA:	4,031	3.2 - 4.8 g/liter
Mouse embryo test (1-cell):	100%	≥ 80%

Product released by Justine Van Renterghem on 14/03/2017 (DD/MM/YYYY)

