

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP17FL11
Production date:	DD/MM/YYYY	14/03/2017
Expiry date:	DD/MM/YYYY	30/09/2018
Total volume:		55000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,38	7.30-7.60
Osmolality:	284	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	0,234 EU/ml	< 0,25EU/ml
Concentration HSA:	4,015	3.2 - 4.8 g/liter
Mouse embryo test (1-cell):	100%	≥ 80%

Product released by Justine Van Renterghem on 30/03/2017 (DD/MM/YYYY)

