

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP17FL16PRG
Production date:	DD/MM/YYYY	12/04/2017
Expiry date:	DD/MM/YYYY	31/10/2018
Total volume:		40000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,42	7.30-7.60
Osmolality:	275	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	<0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	3,293	3.2 - 4.8 g/liter
Concentration gentamicin:	6,72	5 - 15 mg/liter
Mouse embryo test (1-cell):	90,48%	≥ 80%

Product released by Justine Van Renterghem on 27/04/2017 (DD/MM/YYYY)

